

Rotem Industrial Park
Mishor Yamin M.Post Arava 86800 Israel
Tel. -972 8 6570303 Fax. +972 8 6552565
Mail: safety-csr@rotemsafety.co.il



רותם בטיחות בע"מ
Rotem Health&Safety

פארק תעשיות רותם
מישור ימין ד.ג. ערבה 86800
טל. 08-6552565 פקס. 08-6570303
דוא"ל: safety-csr@rotemsafety.co.il

דו"ח כיול RADECO - AIR VOLUME TOTALIZER

CUSTOMER	1023
----------	------

DATE OF ARRIVAL	20/03/2018
ROUT CARD No.	296/18

MODEL No.	H-811	P.O.NO.	106200-2
CAL PROCED. USED:	1006000	SERIAL No.	1136
BAROMETRIC PRESS.	-----	REV.	C
TEMPRETURE	-----	CORRECTED TO	-----
FILTER PAPER USED:	0750-37	CARTRIDGE USED	0750-65
FLOW ELEMENT	C-828	SERIAL No.	1974
(flow element consist of mass flowmeter and laminator flow element)			
HASTINGS POWER SUPPLY	BK PRECISION 1694	SERIAL No.	258G14137

REFERENCE (LPM)	CALIBRATION DATA AS LEFT (LPM)	CALIBRATION DATA AS FOUND
30	27	-----
40	40	-----
50	50	-----
60	60	-----
70	70	-----
80	78	-----

CALIBRATION DATE	22/03/2018
------------------	------------

NEXT CALIBRATION	22/03/2019
------------------	------------

חתימה

אודי מירון

שם המכיל